



1 Remove the mask from the packaging and recycle or discard the plastic insert.



2 Adjust tubing at the swivel to relieve any twisting.




3 Securely fasten connector on oxygen tube to oxygen supply. Ensure all connections are secure and tubing is not kinked.


4

Oxygen flow (LPM)	FiO ₂ (%)
1	24 - 34
2	30 - 50
4	37 - 55
6	47 - 70
8	50 - 79
10	54 - 80
12 - 15	59 - 83


4 Adjust the oxygen flow from 1 LPM to 15 LPM to get the prescribed patient oxygen saturation and oxygen flow rate.




5 Check for oxygen flow through the device.




6 Place device on patient's face. Center the diffuser between the nose and upper lip.




7 Position elastic band below the ears or at the crown of the head, pulling both ends of band to secure.



8 Attach Microstream™ capnography sampling line quick-seal connector to a Microstream™-enabled monitor.



9 Allow the patient to stabilize with oxygen flow, then assess.



10 Adjust the oxygen flow rate as necessary to achieve or maintain prescribed oxygen saturation.

Microstream™ monitoring should not be used as the sole basis for diagnosis or therapy and is intended only as an adjunct in patient assessment.

Review the full product IFU before use by appropriate personnel.

Oxy₂Mask™ is manufactured by Southmedic Inc., and distributed by Medtronic.

©2023 Medtronic. Medtronic, Medtronic logo, and Engineering the extraordinary are trademarks of Medtronic.™ Third party brands are trademarks of their respective owners. All other brands are trademarks of a Medtronic company. 10/2023-US-PM-2300469