


# AVALUS™ BIOPROSTHESIS




## Indications for Use

The Avalus bioprosthesis is indicated for the replacement of a diseased, damaged, or malfunctioning native or prosthetic aortic valve.

## Ordering Information

Avalus Valve Order Number	Valve Size	Stent Diameter (TAD)	Internal Orifice Diameter	External Sewing Ring Diameter	Valve Profile Height	Aortic Protrusion
		(1)	(2)	(3)	(4)	(5)
40019	19 mm	19 mm	17.5 mm	27.0 mm	13.0 mm	11.0 mm
40021	21 mm	21 mm	19.5 mm	29.0 mm	14.0 mm	12.0 mm
40023	23 mm	23 mm	21.5 mm	31.0 mm	15.0 mm	13.0 mm
40025	25 mm	25 mm	23.5 mm	33.0 mm	16.0 mm	14.0 mm
40027	27 mm	27 mm	25.5 mm	36.0 mm	17.0 mm	15.0 mm

TAD – Tissue Annulus Diameter



## Accessories

Order Number	Description
7420	Valve Handle
7400S	Avalus Sizers
T7400	Tray, Accessory, Avalus

INTERNATIONAL

CAUTION: For distribution only in markets where Avalus bioprosthesis has been approved.  
Not for distribution in the U.S.

**Medtronic**  
Further, Together

## Product Specifications

Sterilization	Liquid Chemical Sterilization
Shelf-Life	1.5 years
Packaging (Sterile Barriers)	Valve Jar: Glass Valve Jar Lid: Polypropylene Valve Lid Liner: Silicone
Tissue Fixation	Uniaxial Fixation
Storage Temperature	5°C to 25°C (41°F to 77°F)
Storage Solution	Buffered 0.2% Glutaraldehyde Solution
Rinsing Procedure	2 rinse basins each containing 500mL of sterile, normal saline solution Single 30-second rinse. Store in a second basin until use.
MRI Compatibility	Non Metallic - MR Safe <span style="border: 1px solid black; padding: 0 2px;">MR</span> - Poses no known hazards in all MR environments.

## Materials List

Valve Holder	Blue Acetal Homopolymer
Valve to Valve Holder Sutures	Black Nylon
Valve Leaflets	Bovine Pericardium Cross-linked in 0.5% glutaraldehyde
Wireform	Polyetheretherketone (PEEK)
Base Frame	Polyetheretherketone (PEEK) impregnated with barium sulfate
Fabric Covering Wireform and Base Frame	Polyester
Valve Sewing Cuff	Polyester, Valve Component Sutures, Polyester, Force fiber suture, UHMWPE
Retainer Jar & Retainer cap	Homopolymer
Serial Number Tag	Polypropylene

## Accessories

Avalus Sizer	Polysulfone (handle and head), Nitinol (wire)
Tray	Polyphenylsulfone
Handle	Stainless Steel

## Tissue Treatment

Anticalcification Treatment	Alpha-Amino Oleic Acid (AOA™) treatment
-----------------------------	---

For a listing of indications, contraindications, precautions, warnings, and potential adverse events, please refer to the Instructions for Use. For countries that use eIFUs, consult instructions for use at [www.medtronic.com/manuals](http://www.medtronic.com/manuals). Note: Manuals can be viewed using a current version of any major internet browser. For best results, use Adobe Acrobat® Reader with the browser.

### INTERNATIONAL

CAUTION: For distribution only in markets where Avalus bioprosthesis has been approved. Not for distribution in the U.S.

## Medtronic

710 Medtronic Parkway  
Minneapolis, MN 55432-5604  
USA  
Tel: (763) 514-4000  
Fax: (763) 514-4879  
Toll-free: (800) 328-2518

### LifeLine

**CardioVascular Technical Support**  
Tel: (877) 526-7890  
Tel: (763) 526-7890  
Fax: (763) 526-7888  
[rs.stechsupport@medtronic.com](mailto:rs.stechsupport@medtronic.com)

### Europe

Medtronic International Trading Sàrl  
Route du Molliau 31  
Case Postale  
CH-1131 Tolochenaz  
Switzerland  
Tel: (41 21) 802-7000

### Canada

Medtronic of Canada Ltd.  
99 Hereford Street  
Brampton, Ontario L6Y 0R3  
Canada  
Tel: (905) 460-3800  
Toll-free: (800) 268-5346

### Asia

Medtronic International Ltd.  
49 Changi South Avenue 2  
Nasaco Tech Centre  
Singapore 486056  
Singapore  
Tel: (65) 6436-5000

### Latin America

Doral Corporate Centre II  
3750 NW 87th Avenue, Suite 700  
Miami, FL 33178  
USA  
Tel: (305) 500-9328

©2016 Medtronic.  
All rights reserved.  
Medtronic, Medtronic logo  
and Further Together are  
trademarks of Medtronic.  
Third party brands are trademarks  
of their respective owners.  
All other brands are trademarks  
of a Medtronic company.