

PRODUCT PERFORMANCE REPORT

SUMMARY OF DATA
FROM THE MEDTRONIC
POST-MARKET
REGISTRY

2022

v.1.0 23Mar2023

Medtronic
Further, Together

Contents

5	Deep Brain Stimulation Systems	3
5.1	Study Participants	3
5.1.1	Centers	3
5.1.2	Patients	3
5.2	Event Summary	5
5.2.1	Product Performance Events	5
5.2.2	Clinical Events Not Related To Product Performance	8
5.2.3	Patient Deaths	9
5.3	Neurostimulators	10
5.3.1	Neurostimulator Events	10
5.3.2	Neurostimulator Models	12
5.3.2.1	Model Activa PC	13
5.3.2.2	Model Activa SC	15
5.3.2.3	Model Activa RC	17
5.3.2.4	Model Percept PC	19
5.3.3	Neurostimulator Summary	20
5.4	Leads	21
5.4.1	Lead Events	21
5.4.2	Lead Models	22
5.4.2.1	Model 3387	23
5.4.2.2	Model 3389	26
5.4.2.3	Model 3391	29
5.4.3	Lead Summary	30
5.5	Extensions	31
5.5.1	Extension Events	32
5.5.2	Extension Models	32
5.5.2.1	Model 37085/37086	33
5.5.2.2	Model B34000/B34000M	36
5.5.3	Extension Summary	37

5 Deep Brain Stimulation Systems

5.1 Study Participants

5.1.1 Centers

In this section, the deep brain stimulation tables and graphs were generated based on data collected between July 2009 and the report cut-off date of October 31, 2022. Sixty-four centers in North America, Europe, South America, Asia, and Australia have enrolled and contributed patients to the deep brain stimulation systems section of this report. [Figure 5.1](#) shows a World Map, in which the countries that enrolled DBS patients are highlighted.




Figure 5.1: Countries with Deep Brain Stimulation Therapy Patients in Registry (Highlighted)

5.1.2 Patients

Of the 3,295 deep brain stimulation patients enrolled, the primary indications for implant were as follows: 61.1% were implanted for the treatment of Parkinson's Disease, 22.6% were implanted for the treatment of essential tremor, 9.9% were implanted for the treatment of dystonia, 1.5% were implanted for the treatment of epilepsy, 1.5% were implanted for the treatment of obsessive compulsive disorder, 2.3% were implanted for the treatment of other indications, and 1.0% were implanted for indications that were not specified in the database at the time of data cut-off (see [Figure 5.2](#) and [Table 5.1](#)).




Figure 5.2: Deep Brain Stimulation Primary Treatment Indications

Table 5.1: Deep Brain Stimulation Primary Treatment Indications

Primary Treatment Indication ^a	Enrolled Patients (%)
Parkinson's Disease	2,014 (61.1%)
Essential Tremor	746 (22.6%)
Dystonia	327 (9.9%)
Epilepsy	50 (1.5%)
OCD	49 (1.5%)
Other	77 (2.3%)
Not Specified	32 (1.0%)
Total Patients	3,295(100%)

^a For approved indications refer to product labeling for your geography.

It is recognized that healthcare providers prescribe therapies to meet specific patient needs; however, Medtronic only directs the use of its products based on approved labeling, which varies by geography. Please contact your local Medtronic representative for region-specific product labeling (<http://www.medtronic.com/us-en/about/locations.html>).

5.2 Event Summary

There were 479 product performance events reported between July 2009 and October 31, 2022, in patients with deep brain stimulation systems. These events represent 22.6% of the total reported events (479/2,124), occurred in 294 of the 3,295 (8.9%) total patients enrolled, and are presented graphically within this report (e.g. events per patient years as well as survival curves). Of the remaining 1,645 reported events 269 were serious (not product performance related) and 1,376 were non-serious (not product performance related). Serious non-product performance related events (n=269) are described in [Table 5.6](#). Non serious non-product performance related (n=1,376) events are not listed in this report.

Any registry devices that are returned to Medtronic are analyzed via a Returned Product Analysis (RPA) process. If available, RPA findings overwrite in the classification of the events. Within this report, [Table 5.2](#) differentiates the events by those determined by the RPA process versus those determined by the physician. Please refer to the Methodology section for more information.

There were 385 deaths reported for patients followed in the PSR with deep brain stimulation systems (see [Table 5.7](#)), none of which were reported as a direct result of a product performance event. Early versions of the protocol required events to be reported only when the event required a surgical intervention, resulted in therapy abandonment, or resulted in death. The required event reporting definition was expanded in April 2010 to include all adverse events related to the device, implant procedure, and/or therapy.

5.2.1 Product Performance Events

Table 5.2: Deep Brain Stimulation System Product Performance Events

Product Performance Events ^a	Event Counts	Events Per 100 Patient Years	Patients with Events (%) N=3,295 ^b
RPA Determination	3	0.03	3 (0.09%)
Premature Battery Depletion	3	0.03	3 (0.09%)
Physician's Determination	476	4.08	292 (8.86%)
High Impedance	215	1.84	131 (3.98%)
Lead Migration/Dislodgement	46	0.39	33 (1.00%)
Low Impedance	32	0.27	21 (0.64%)
Device Malfunction	24	0.21	20 (0.61%)
Lead Fracture	22	0.19	18 (0.55%)
Extension Migration	21	0.18	11 (0.33%)
Neurostimulator Unable To Recharge ^c	18	0.15	18 (0.55%)
Extension Fracture	12	0.1	9 (0.27%)
Premature Battery Depletion	10	0.09	10 (0.30%)
Medical Device Complication	9	0.08	6 (0.18%)
Device Breakage	5	0.04	5 (0.15%)
Device Electrical Finding ^d	5	0.04	4 (0.12%)

...continued

Product Performance Events^a	Event Counts	Events Per 100 Patient Years	Patients with Events (%) N=3,295^b
Medical Device Site Infection	5	0.04	4 (0.12%)
Device Lead Issue	4	0.03	4 (0.12%)
Device Protrusion	4	0.03	2 (0.06%)
Device Connection Issue	3	0.03	3 (0.09%)
Device Electrical Impedance Issue	3	0.03	2 (0.06%)
Device End Of Life	3	0.03	3 (0.09%)
Electric Shock Sensation	3	0.03	2 (0.06%)
Neurostimulator Inversion	3	0.03	3 (0.09%)
Device Charging Issue	2	0.02	2 (0.06%)
Device Material Corroded	2	0.02	1 (0.03%)
Device Short Circuiting	2	0.02	2 (0.06%)
Electromagnetic Interference	2	0.02	2 (0.06%)
Medical Device Site Erosion	2	0.02	2 (0.06%)
Other ^e	19	0.16	17 (0.52%)
Total	479	4.11	294 (8.92%)

^a Medical Dictionary for Regulatory Activities (MedDRA) Lower-Level Term or Medtronic's coding system term for events that do not exist in the MedDRA dictionary.

^b The total number of patients with events may not represent the sum of all rows, as a patient may have experienced more than one type of event.

^c There were 721 patients that used rechargeable neurostimulators for DBS in the registry. A total of 2.5% (18/721) of patients with a rechargeable neurostimulator experienced a neurostimulator unable to recharge event.

^d Including open circuit contact, electric discharge.

^e Composed of event codes with 1 event each.

A total of 221 (46.1%) of the 479 product performance events were related to the lead, 92 (19.2%) were related to the extension, 84 (17.5%) were related to the neurostimulator, 15 (3.1%) were related to multiple etiologies, which includes events where at least one device and one non-device etiology was indicated, and 67 (14.0%) were related to other etiologies, including: 36 (7.5%) were related to other component, 10 (2.1%) were related to incisional site/device tract, 9 (1.9%) were related to surgery/anesthesia, 6 (1.3%) were related to recharging process, 5 (1.0%) were related to programming/stimulation, and 1 (0.2%) was related to other etiology (see [Figure 5.3](#)). Events could have more than one etiology.

Relatedness is reported by the physician. In cases where the Clinical Events Committee (CEC) has adjudicated relatedness differently from the site, the CEC adjudication is used in this report for analysis purposes. However, both the site's reporting and the CEC's adjudication remain in the database.




Figure 5.3: Deep Brain Stimulation System Product Performance Events by Relatedness

Table 5.3 and Table 5.4 describe the interventions taken for reported impedance events. In 39.1% and 31.2% of the high and low impedance events, the action taken was a surgical intervention. However, impedance could be used as a diagnostic measurement and may not result in any intervention or clinical impact. The majority of events required no intervention, or device reprogramming only (54.9% for high impedance and 68.8% for low impedance).

Table 5.3: Deep Brain Stimulation System High Impedance Events by Intervention

Intervention	N (%) of High Impedance Events
Device Surgical Intervention	72 (33.5%)
Reprogramming	52 (24.2%)
Other Surgical Intervention	12 (5.6%)
Other Intervention	7 (3.3%)
Medical or Non-Surgical Therapy	5 (2.3%)
Therapy Suspension	1 (0.5%)
No Action Taken	66 (30.7%)
Total	215 (100%)

Table 5.4: Deep Brain Stimulation System Low Impedance Events by Intervention

Intervention	N (%) of Low Impedance Events
Device Surgical Intervention	10 (31.2%)
Reprogramming	10 (31.2%)
No Action Taken	12 (37.5%)
Total	32 (100%)

Table 5.5 describes the interventions taken for reported lead migration/dislodgement events; 78.3% of them led to a surgical intervention, and 13.0% were reprogramming.

Table 5.5: Deep Brain Stimulation System Lead Migration/Dislodgement Events by Intervention

Intervention	N (%) of Lead Migration/Dislodgement Events
Device Surgical Intervention	34 (73.9%)
Reprogramming	6 (13.0%)
Other Surgical Intervention	2 (4.3%)
Medication	1 (2.2%)
No Action Taken	3 (6.5%)
Total	46 (100%)

5.2.2 Clinical Events Not Related To Product Performance

The clinical events not related to product performance are summarized if:

- The patient was enrolled in the PSR at the time in which the clinical event collection was initiated (N=2,112)
- Categorized as serious adverse events (SAEs, N=269)
- Occurred with a System Organ Class (SOC) threshold $\geq 1\%$ of patients
- Other Considerations
 - Some events are described in high level group terms (HLGT) to provide more specificity, if needed
 - Some therapies will provide therapy relevant events

Table 5.6: Deep Brain Stimulation System Clinical Events Not Related To Product Performance

Event Type	Number of SAE	Patients with SAE n (%) ^a N=2,112	SAE Per 100 Patient Months	Patient with SAE Requiring Surgical Intervention n (%) N=2,112
Infections and infestations	107	87 (4.12%)	0.13	81 (3.84%)
Infections - pathogen unspecified	104	84 (3.98%)	0.13	78 (3.69%)
Other ^b	3	3 (0.14%)	0.00	3 (0.14%)
Nervous system disorders	59	52 (2.46%)	0.07	11 (0.52%)
Central nervous system vascular disorders	19	19 (0.90%)	0.02	3 (0.14%)
Movement disorders (incl parkinsonism)	18	18 (0.85%)	0.02	3 (0.14%)
Neurological disorders NEC	11	10 (0.47%)	0.01	3 (0.14%)
Other ^b	11	11 (0.52%)	0.01	2 (0.09%)
General disorders and administration site conditions	49	46 (2.18%)	0.06	29 (1.37%)
Complications associated with device	36	34 (1.61%)	0.05	25 (1.18%)
General system disorders NEC	7	7 (0.33%)	0.01	1 (0.05%)
Other ^b	6	6 (0.28%)	0.01	3 (0.14%)
Other SOC Terms (<1.0% Threshold)	54	52 (2.46%)	0.07	19 (0.90%)
Total	269	212 (10.04%)	0.34	130 (6.16%)

^a The total number of patients with events may not represent the sum of all rows, as a patient may have experienced more than one type of event.

^b Composed of high level group term event codes with fewer than 5 events each.

5.2.3 Patient Deaths

There were 385 deaths reported for patients with deep brain stimulation systems, none of which were reported as a direct result of a product performance event. Since July 2009, a total of 310 (80.5%) deaths have been reported in this patient registry study based upon patients receiving therapy for Parkinson's Disease, 51 (13.2%) for essential tremor, 15 (3.9%) for dystonia, 3 (0.8%) for epilepsy, 1 (0.3%) for OCD, and 5 (1.3%) for other indications (see [Table 5.7](#)). The percentage is based upon the total patient death events and not based upon the rate of occurrence.

Table 5.7: Deep Brain Stimulation System Patient Deaths by Primary Indication

Number of Reports of Death by Primary Indication ^a	N (%) of Deaths	Mean Age of Death in Years
Parkinson's Disease	310 (80.5%)	74.06
Essential Tremor	51 (13.2%)	79.8
Dystonia	15 (3.9%)	56.1
Epilepsy	3 (0.8%)	37.6
OCD	1 (0.3%)	70.4
Other	5 (1.3%)	75.3
Total	385 (100%)	74.1

^a For approved indications refer to product labeling for your geography.

5.3 Neurostimulators

From July 2009 to the report cut-off date of October 31, 2022, there were 5,212 neurostimulators followed in the registry. The difference between the total number of patients (n=3,295) versus the number of neurostimulators (n=5,212) is due to the fact that some patients were implanted with more than one neurostimulator or were subsequently re-implanted. The aggregate prospective follow-up time for all neurostimulators was 150,157 months (12,513 years). [Table 5.8](#) provides the number and percentage of neurostimulators by model.

Table 5.8: Deep Brain Stimulation Neurostimulator Counts by Model

Model Name	N (%)
Currently manufactured	
Activa PC	2,730 (52.4%)
Activa SC	1,106 (21.2%)
Activa RC	738 (14.2%)
Percept PC	526 (10.1%)
No longer manufactured	
Soletra	67 (1.3%)
Other/Unspecified ^a	33 (0.6%)
Total	5,212 (100%)

^a Other includes Activa PC+S and non-Activa systems used for DBS.

5.3.1 Neurostimulator Events

Of the total of 479 product performance-related events, there were 86 product performance-related events with an underlying reported etiology related to neurostimulator function. This includes 84 events with a neurostimulator etiology and 2 events with both a neurostimulator and other etiology (including device and non-device etiologies). Of these, 66 were the initial product performance events that affected neurostimulator survival estimates. For neurostimulators in the registry, the current return rate to Medtronic Returned Product Analysis (RPA) was 4.1% (94/2,311). The proportion was based upon the number of registry neurostimulators received by RPA, divided by the sum of the total number of explanted devices and the total number of neurostimulators in patients who have expired. In the 86 neurostimulator events, 96.5 % (83/86) were assigned as device related by the physician, not returned to Medtronic RPA (see [Table 5.9](#)).

Table 5.9: Deep Brain Stimulation Neurostimulator Product Performance Events by Determination

Product Performance Events	N (%)
RPA Determination	3 (3.5%)
Premature Battery Depletion	3 (3.5%)
Physician's Determination	83 (96.5%)
High Impedance	37 (43.0%)
Device Malfunction	10 (11.6%)
Premature Battery Depletion	10 (11.6%)
Low Impedance	6 (7.0%)
Device End Of Life	3 (3.5%)
Device Electrical Finding ^a	2 (2.3%)
Electromagnetic Interference	2 (2.3%)
Extension Migration	2 (2.3%)
Device Computer Software Issue	1 (1.2%)
Device Protrusion	1 (1.2%)
Device Short Circuiting	1 (1.2%)
Device Vibration	1 (1.2%)
Electric Shock Sensation	1 (1.2%)
Medical Device Site Discomfort	1 (1.2%)
Medical Device Site Erosion	1 (1.2%)
Medical Device Site Infection	1 (1.2%)
Neurostimulator Inversion	1 (1.2%)
Neurostimulator Unable To Recharge	1 (1.2%)
Wound Infection	1 (1.2%)
Total	86 (100%)

^a Open circuit contact.

For the purposes of survival analysis, a device's follow-up time is cut-off for one of three reasons: 1) the occurrence of a product performance-related event; 2) the occurrence of a censoring event; or 3) the device is event-free and censored at the patient's last follow-up prior to the data cut-off. For neurostimulators:

- 66 had follow-up time cut-off due to product performance-related events.
- 3,390 were censored in the survival analysis for the following reasons: patient expired, neurostimulator explanted, patient discontinued, therapy suspended, or site discontinued participation in the registry.
- 1,756 were free from product performance-related events and censoring events, and were censored at the last follow-up visit prior to the report cut-off.


5.3.2 Neurostimulator Models

The following figures and tables represent neurostimulator survival and 95% confidence intervals where at least 20 neurostimulators contributed to each 3-month interval.

The Soletra and Kinetra models were removed from the table due to the limited number of active devices in PSR. For information on survival for those models, please refer to past reports.

5.3.2.1 Model Activa PC

Model Name	Activa PC
FDA Approval Date	April 2009
Neurostimulators Enrolled	2,730
Neurostimulators Currently Active in Study	584
Initial Product Performance Events	42
Median Follow-up Time (Months)	29.2
Cumulative Follow-up Time (Months)	87,874



Time Interval	1 Year	2 Years	3 Years	4 Years	5 Years
Survival	99.2%	98.8%	98.4%	98.0%	97.8%
(95% CI)	(98.7%, 99.5%)	(98.2%, 99.2%)	(97.8%, 98.9%)	(97.2%, 98.6%)	(96.8%, 98.5%)
Sample Size	2,193	1,640	1,042	600	323

Time Interval	6 Years	7 Years	8 Years	At 99 Months	
Survival	97.1%	97.1%	97.1%	97.1%	
(95% CI)	(95.1%, 98.3%)	(95.1%, 98.3%)	(95.1%, 98.3%)	(95.1%, 98.3%)	—
Sample Size	125	60	23	21	

Specification: Activa PC	
Height	2.6 in (65 mm)
Width	1.9 in (49 mm)
Thickness	0.6 in (15 mm)
Volume	39 cc
Battery type	Non-Rechargeable
Expected Battery life	Depends on settings and use
Maximum Electrodes	8
Amplitude	0 - 10.5 V (voltage mode) 0 - 25.5 mA (current mode)
Rate	2 - 250 Hz (voltage mode) 30 - 250 Hz (current mode)
Pulse Width	60 - 450 µsec
Groups	4
Programs	16 (up to 4 per group)
Implant Depth	≤ 4 cm




Neurostimulator Event: Activa PC	N
High impedance	17
Premature battery depletion	12
Device malfunction	5
Device end of life	3
Device electrical finding ^a	1
Device protrusion	1
Electromagnetic interference	1
Low impedance	1
Medical device site infection	1
Total	42

^a Open circuit contact.

5.3.2.2 Model Activa SC

Model Name	Activa SC
FDA Approval Date	January 2011
Neurostimulators Enrolled	1,106
Neurostimulators Currently Active in Study	206
Initial Product Performance Events	9
Median Follow-up Time (Months)	26.5
Cumulative Follow-up Time (Months)	31,524



Time Interval	1 Year	2 Years	3 Years	4 Years	5 Years
Survival	99.6%	99.5%	99.3%	98.7%	98.7%
(95% CI)	(98.9%, 99.8%)	(98.7%, 99.8%)	(98.4%, 99.7%)	(97.1%, 99.4%)	(97.1%, 99.4%)
Sample Size	846	613	359	183	77

Time Interval	6 Years	At 78 Months			
Survival	98.7%	98.7%	—	—	—
(95% CI)	(97.1%, 99.4%)	(97.1%, 99.4%)	—	—	—
Sample Size	38	27			


Specification: Activa SC	
Height	2.2 in (55 mm)
Width	2.4 in (60 mm)
Thickness	0.4 in (11 mm)
Volume	28 cc (Model 37602) 27 cc (Model 37603)
Battery type	Non-Rechargeable
Expected Battery life	Depends on settings and use
Maximum Electrodes	4
Amplitude	0 - 10.5 V (voltage mode) 0 - 25.5 mA (current mode)
Rate	2 - 250 Hz (voltage mode) 30 - 250 Hz (current mode)
Pulse Width	60 - 450 μ sec
Groups	4
Programs	8 (up to 2 per group)
Implant Depth	\leq 4 cm



Neurostimulator Event: Activa SC	N
High impedance	4
Device short circuiting	1
Low impedance	1
Medical device site discomfort	1
Premature battery depletion	1
Wound infection	1
Total	9

5.3.2.3 Model Activa RC

Model Name	Activa RC
FDA Approval Date	March 2009
Neurostimulators Enrolled	738
Neurostimulators Currently Active in Study	489
Initial Product Performance Events	12
Median Follow-up Time (Months)	24.8
Cumulative Follow-up Time (Months)	24,295



Time Interval	1 Year	2 Years	3 Years	4 Years	5 Years
Survival	99.5%	99.1%	98.8%	98.0%	96.7%
(95% CI)	(98.5%, 99.8%)	(97.7%, 99.6%)	(97.3%, 99.5%)	(95.8%, 99.0%)	(93.3%, 98.3%)
Sample Size	523	378	269	196	138

Time Interval	6 Years	7 Years	8 Years	At 105 Months	
Survival	96.7%	96.7%	96.7%	96.7%	
(95% CI)	(93.3%, 98.3%)	(93.3%, 98.3%)	(93.3%, 98.3%)	(93.3%, 98.3%)	—
Sample Size	86	52	29	21	


Specification: Activa RC	
Height	2.1 in (54 mm)
Width	2.1 in (54 mm)
Thickness	0.4 in (9 mm)
Volume	22 cc
Battery type	Rechargeable
Expected Battery life	9 years
Maximum Electrodes	8
Amplitude	0 - 10.5 V (voltage mode) 0 - 25.5 mA (current mode)
Rate	2 - 250 Hz (voltage mode) 30 - 250 Hz (current mode)
Pulse Width	60 - 450 µsec
Groups	4
Programs	16 (up to 4 per group)
Implant Depth	≤ 1 cm



Neurostimulator Event: Activa RC	N
Device malfunction	2
Extension migration	2
High impedance	2
Device computer software issue	1
Electric shock sensation	1
Low impedance	1
Medical device site erosion	1
Neurostimulator inversion	1
Neurostimulator unable to recharge	1
Total	12


5.3.2.4 Model Percept PC

Model Name	Percept PC
FDA Approval Date	June 2020
Neurostimulators Enrolled	526
Neurostimulators Currently Active in Study	477
Initial Product Performance Events	2
Median Follow-up Time (Months)	4.9
Cumulative Follow-up Time (Months)	3,306



Time Interval	1 Year	At 18 Months
Survival	99.3%	99.3%
(95% CI)	(97.1%, 99.8%)	(97.1%, 99.8%)
Sample Size	110	36

Specification: Percept PC	
Height	68 mm
Width	2 in (51 mm)
Thickness	0.43 in (11 mm)
Volume	33 cc
Battery type	Non-Rechargeable
Expected Battery life	Depends on settings and use
Maximum Electrodes	16 Electrodes (8 per lead)
Amplitude	0 to 25.5 mA
Rate	2 to 250 Hz
Pulse Width	20 to 450 µsec
Groups	4
Programs	16
Implant Depth	≤ 4cm



Neurostimulator Event: Percept PC	N
Device electrical finding ^a	1
High impedance	1
Total	2

^a Open circuit contact.

5.3.3 Neurostimulator Summary

Table 5.10: Deep Brain Stimulation Neurostimulator Characteristics

Model Name	FDA Approval Date	Neurostimulators Enrolled	Neurostimulators Active	Initial Product Performance Events	Median Follow-up Time (Months)	Cumulative Follow-up Time (Months)
Activa PC	April 2009	2,730	584	42	29.2	87,874
Activa SC	January 2011	1,106	206	9	26.5	31,524
Activa RC	March 2009	738	489	12	24.8	24,295

Table 5.11: Deep Brain Stimulation Neurostimulator Survival Probability (95% Confidence Intervals)

Model Name	1 Year	2 Years	3 Years	4 Years	5 Years	6 Years	7 Years	8 Years
Activa PC	99.2% (98.7%, 99.5%)	98.8% (98.2%, 99.2%)	98.4% (97.8%, 98.9%)	98.0% (97.2%, 98.6%)	97.8% (96.8%, 98.5%)	97.1% (95.1%, 98.3%)	97.1% (95.1%, 98.3%)	97.1% (95.1%, 98.3%)
Activa SC	99.6% (98.9%, 99.8%)	99.5% (98.7%, 99.8%)	99.3% (98.4%, 99.7%)	98.7% (97.1%, 99.4%)	98.7% (97.1%, 99.4%)	98.7% (97.1%, 99.4%)	—	—
Activa RC	99.5% (98.5%, 99.8%)	99.1% (97.7%, 99.6%)	98.8% (97.3%, 99.5%)	98.0% (95.8%, 99.0%)	96.7% (93.3%, 98.3%)	96.7% (93.3%, 98.3%)	96.7% (93.3%, 98.3%)	96.7% (93.3%, 98.3%)

5.4 Leads

From July 2009 to the report cut-off date of October 31, 2022, there were 5,602 leads followed in the registry. The difference between the total number of leads (n=5,602) versus neurostimulators (n=5,212) is due to the fact that some patients were subsequently re-implanted with a lead or were implanted with more than one lead. The aggregate prospective follow-up time for all leads was 241,313 months (20,109 years). [Table 5.12](#) provides the number and percentage of leads by model.

Table 5.12: Deep Brain Stimulation Lead Counts by Model

Model Name	N (%)
3389 (compact electrode spacing)	2,935 (52.4%)
3387 (standard electrode spacing)	2,269 (40.5%)
SenSight B33005 (compact electrode spacing)	237 (4.2%)
SenSight B33015 (standard electrode spacing)	73 (1.3%)
3391 (large electrodes and wide spacing)	68 (1.2%)
Other/Unspecified ^a	20 (0.4%)
Total	5,602 (100%)

^a Includes leads used in non-Activa systems.

5.4.1 Lead Events

Of the total of 479 product performance-related events, there were 230 product performance-related events with an underlying reported etiology related to lead function. This includes 221 events with a lead etiology and 9 events with both a lead and other etiology (including device and non-device etiologies). Of these, 134 were the initial product performance event that affected lead survival estimates.

Events of other/unspecified models are not shown. Model 3391 did not have any product performance-related events.

For the purposes of survival analysis, a device's follow-up time is cut-off for one of three reasons: 1) the occurrence of a product performance-related event; 2) the occurrence of a censoring event; or 3) the device is event-free and censored at the patient's last follow-up prior to the data cut-off. For leads:


- 134 had follow-up time cut-off due to product performance-related events.
- 2,547 were censored in the survival analysis for the following reasons: patient expired, neurostimulator explanted, patient discontinued, therapy suspended, or site discontinued participation in the registry.
- 2,921 were free from product performance-related events and censoring events, and were censored at the last follow-up visit prior to the report cut-off.

5.4.2 Lead Models

The following figures and tables represent lead survival and 95% confidence intervals where at least 20 leads contributed to each 3-month interval. Due to enrollment of replacement patients with previously implanted leads, sample size may increase at later timepoints.

5.4.2.1 Model 3387

Model Name	3387
FDA Approval Date	July 1997
Leads Enrolled	2,269
Leads Currently Active in Study	1,144
Initial Product Performance Events	43
Median Follow-up Time (Months)	36.3
Cumulative Follow-up Time (Months)	97,797



Time Interval	1 Year	2 Years	3 Years	4 Years	5 Years
Survival (95% CI)	99.0% (98.4%, 99.4%)	98.9% (98.2%, 99.3%)	98.5% (97.7%, 99.0%)	98.0% (97.1%, 98.7%)	97.5% (96.5%, 98.3%)
Sample Size	1,532	1,189	1,000	851	714
Time Interval	6 Years	7 Years	8 Years	9 Years	10 Years
Survival (95% CI)	97.4% (96.3%, 98.2%)	97.2% (96.0%, 98.0%)	96.6% (95.1%, 97.7%)	96.6% (95.1%, 97.7%)	96.1% (94.1%, 97.4%)
Sample Size	512	354	253	187	140
Time Interval	11 Years	12 Years	13 Years	14 Years	15 Years
Survival (95% CI)	95.3% (92.3%, 97.1%)	95.3% (92.3%, 97.1%)	93.7% (88.5%, 96.6%)	90.2% (82.1%, 94.7%)	90.2% (82.1%, 94.7%)
Sample Size	108	62	56	53	35
Time Interval	16 Years	At 198 Months			
Survival (95% CI)	90.2% (82.1%, 94.7%)	90.2% (82.1%, 94.7%)	—	—	—
Sample Size	24	22			

Specification: 3387	
Lead	
Length (cm)	40
Diameter (mm)	1.27
Electrode	
Number	4
Shape	Cylindrical
Length (mm)	1.5
Individual Surface Area (mm ²)	6.0
Inter-Electrode Spacing: Edge to Edge (mm)	1.5
Array Length (mm)	10.5




Lead Event: 3387	N
High impedance	20
Lead migration/dislodgement	10
Low impedance	7
Lead fracture	3
Device electrical finding ^a	1
Device lead issue	1
Medical device site pain	1
Total	43

^a Open circuit contact.

5.4.2.2 Model 3389

Model Name	3389
FDA Approval Date	September 1999
Leads Enrolled	2,935
Leads Currently Active in Study	1,483
Initial Product Performance Events	87
Median Follow-up Time (Months)	44.7
Cumulative Follow-up Time (Months)	138,874



Time Interval	1 Year	2 Years	3 Years	4 Years	5 Years
Survival (95% CI)	98.3% (97.7%, 98.8%)	97.9% (97.1%, 98.4%)	97.3% (96.5%, 98.0%)	96.5% (95.5%, 97.3%)	96.0% (95.0%, 96.9%)
Sample Size	1,879	1,565	1,405	1,257	1,091
Time Interval	6 Years	7 Years	8 Years	9 Years	10 Years
Survival (95% CI)	95.6% (94.5%, 96.5%)	94.9% (93.6%, 96.0%)	93.6% (91.8%, 94.9%)	93.6% (91.8%, 94.9%)	93.2% (91.2%, 94.7%)
Sample Size	818	600	414	295	218
Time Interval	11 Years	12 Years	13 Years	14 Years	15 Years
Survival (95% CI)	92.7% (90.4%, 94.4%)	92.7% (90.4%, 94.4%)	92.7% (90.4%, 94.4%)	92.7% (90.4%, 94.4%)	92.7% (90.4%, 94.4%)
Sample Size	183	152	131	97	74
Time Interval	16 Years	17 Years	At 213 Months		
Survival (95% CI)	91.1% (86.6%, 94.1%)	91.1% (86.6%, 94.1%)	91.1% (86.6%, 94.1%)	—	—
Sample Size	54	36	21		


Specification: 3389	
Lead	
Length (cm)	40
Diameter (mm)	1.27
Electrode	
Number	4
Shape	Cylindrical
Length (mm)	1.5
Individual Surface Area (mm ²)	6.0
Inter-Electrode Spacing: Edge to Edge (mm)	0.5
Array Length (mm)	7.5



Lead Event: 3389	N
High impedance	43
Lead migration/dislodgement	20
Lead fracture	12
Low impedance	4
Device material corroded	2
Medical device complication	2
Medical device site infection	2
Device lead issue	1
Lead insulation failure	1
Total	87

5.4.2.3 Model 3391

Model Name	3391
FDA Approval Date	February 2009
Leads Enrolled	68
Leads Currently Active in Study	48
Initial Product Performance Events	0
Median Follow-up Time (Months)	29.0
Cumulative Follow-up Time (Months)	2,607



Time Interval	1 Year	2 Years	At 27 Months
Survival	100.0%	100.0%	100.0%
(95% CI)	(NA)	(NA)	(NA)
Sample Size	38	26	26

Specification: 3391	
Lead	
Length (cm)	40
Diameter (mm)	1.27
Electrode	
Number	4
Shape	Cylindrical
Length (mm)	3.0
Individual Surface Area (mm ²)	12
Inter-Electrode Spacing: Edge to Edge (mm)	4.0
Array Length (mm)	24



5.4.3 Lead Summary

Table 5.13: Deep Brain Stimulation Lead Characteristics

Model Name	FDA Approval Date	Leads Enrolled	Leads Active	Initial Product Performance Events	Median Follow-up Time (Months)	Cumulative Follow-up Time (Months)
3387	July 1997	2,269	1,144	43	36.3	97,797
3389	September 1999	2,935	1,483	87	44.7	138,874
3391	February 2009	68	48	0	29.0	2,607

Table 5.14: Deep Brain Stimulation Lead Survival Probability (95% Confidence Intervals)

Model Name	1 Year	2 Years	3 Years	4 Years	5 Years
3387	99.0% (98.4%, 99.4%)	98.9% (98.2%, 99.3%)	98.5% (97.7%, 99.0%)	98.0% (97.1%, 98.7%)	97.5% (96.5%, 98.3%)
3389	98.3% (97.7%, 98.8%)	97.9% (97.1%, 98.4%)	97.3% (96.5%, 98.0%)	96.5% (95.5%, 97.3%)	96.0% (95.0%, 96.9%)
3391	100.0% (NA)	100.0% (NA)	—	—	—
Model Name	6 Years	7 Years	8 Years	9 Years	10 Years
3387	97.4% (96.3%, 98.2%)	97.2% (96.0%, 98.0%)	96.6% (95.1%, 97.7%)	96.6% (95.1%, 97.7%)	96.1% (94.1%, 97.4%)
3389	95.6% (94.5%, 96.5%)	94.9% (93.6%, 96.0%)	93.6% (91.8%, 94.9%)	93.6% (91.8%, 94.9%)	93.2% (91.2%, 94.7%)
3391	—	—	—	—	—
Model Name	11 Years	12 Years	13 Years	14 Years	15 Years
3387	95.3% (92.3%, 97.1%)	95.3% (92.3%, 97.1%)	93.7% (88.5%, 96.6%)	90.2% (82.1%, 94.7%)	90.2% (82.1%, 94.7%)
3389	92.7% (90.4%, 94.4%)	92.7% (90.4%, 94.4%)	92.7% (90.4%, 94.4%)	92.7% (90.4%, 94.4%)	92.7% (90.4%, 94.4%)
3391	—	—	—	—	—
Model Name	16 Years	17 Years			
3387	90.2% (82.1%, 94.7%)	—	—	—	—
3389	91.1% (86.6%, 94.1%)	91.1% (86.6%, 94.1%)	—	—	—
3391	—	—	—	—	—

5.5 Extensions

From July 2009 to the report cut-off date of October 31, 2022, there were 5,657 extensions followed in the registry. The difference between the total number of extensions (n=5,657) versus neurostimulators (n=5,212) is due to some patients implanted with more than 1 extension or subsequently re-implanted with an extension. The aggregate prospective follow-up time for all extensions was 236,127 months (19,677 years). [Table 5.15](#) provides the number and percentage of extensions by model.

Table 5.15: Deep Brain Stimulation Extension Counts by Model

Model Name	N (%)
Currently manufactured	
37085/37086 (quadripolar stretch)	4,735 (83.7%)
SenSight B34000/B34000M	309 (5.5%)
No longer manufactured	
7482 ^b (quadripolar)	491 (8.7%)
Other/Unspecified ^a	122 (2.2%)
Total	5,657 (100%)

^a Includes extensions for other legacy stimulation systems.

^b Includes Models 7482 and 7482a.

5.5.1 Extension Events

Of the total of 479 product performance-related events, there were 96 product performance-related events with an underlying reported etiology related to extension function. This includes 92 events with an extension etiology and 4 events with both an extension and other etiology (including device and non-device etiologies). Of these, 82 were the initial product performance event that affected extension survival estimates.

Events of other/unspecified models and discontinued models are not shown.

For the purposes of survival analysis, a device's follow-up time is cut-off for one of three reasons: 1) the occurrence of a product performance-related event; 2) the occurrence of a censoring event; or 3) the device is event-free and censored at the patient's last follow-up prior to the data cut-off. For extensions:


- 82 had follow-up time cut-off due to product performance-related events.
- 2,620 were censored in the survival analysis for the following reasons: patient expired, neurostimulator explanted, patient discontinued, therapy suspended, or site discontinued participation in the registry.
- 2,955 were free from product performance-related events and censoring events, and were censored at the last follow-up visit prior to the report cut-off.

5.5.2 Extension Models

The following figures and tables represent extension survival and 95% confidence intervals where at least 20 extensions contributed to each 3-month interval.

5.5.2.1 Model 37085/37086

Model Name	37085/37086
FDA Approval Date	March 2009/February 2012
Extensions Enrolled	4,735
Extensions Currently Active in Study	2,471
Initial Product Performance Events	68
Median Follow-up Time (Months)	37.6
Cumulative Follow-up Time (Months)	203,009




Time Interval	1 Year	2 Years	3 Years	4 Years	5 Years
Survival (95% CI)	99.2% (98.9%, 99.5%)	98.9% (98.4%, 99.2%)	98.6% (98.2%, 99.0%)	98.5% (98.0%, 98.9%)	98.1% (97.5%, 98.5%)
Sample Size	3,418	2,741	2,362	2,005	1,619

Time Interval	6 Years	7 Years	8 Years	9 Years	10 Years
Survival (95% CI)	98.1% (97.5%, 98.5%)	97.7% (97.0%, 98.3%)	97.7% (97.0%, 98.3%)	97.7% (97.0%, 98.3%)	97.7% (97.0%, 98.3%)
Sample Size	1,143	722	418	241	113

Time Interval	11 Years	At 138 Months			
Survival (95% CI)	94.5% (87.7%, 97.6%)	89.5% (78.1%, 95.1%)	—	—	—
Sample Size	44	24			

Specification: 37085/37086	
Device Name	Stretch-Coil Extension
Length (cm)	40, 60, 95
Distal End Compatibility	3387, 3389, or 3391 lead
Distal End Set Screws	4
Proximal End INS Compatibility	Activa RC, Activa PC, Activa SC, or Percept PC




Extension Event: 37085/37086	Total
High impedance	27
Extension migration	15
Extension fracture	8
Low impedance	4
Medical device complication	4
Device protrusion	3
Device electrical finding ^a	2
Electric shock sensation	2
Device malfunction	1
Dystonia ^b	1
Lead migration/dislodgement	1
Total Extension Events	68

^a Open circuit contact.

^b Device recharging process issue.


5.5.2.2 Model B34000/B34000M

Model Name	B34000/B34000M
FDA Approval Date	May 2021
Extensions Enrolled	309
Extensions Currently Active in Study	290
Initial Product Performance Events	2
Median Follow-up Time (Months)	1.2
Cumulative Follow-up Time (Months)	922



Time Interval	6 Months	9 Months
Survival	97.5%	97.5%
(95% CI)	(90.3%, 99.4%)	(90.3%, 99.4%)
Sample Size	62	33

Specification: B34000/B34000M	
Device Name	SenSight Extension Kit
Length (cm)	40, 60, 95
Distal End Compatibility	B33005, or B33015 Lead
Distal End Set Screws	1
Proximal End INS Compatibility	Activa RC, Percept PC , or Percept RC



Extension Event: B34000/B34000M	Total
Extension migration	2
Total Extension Events	2

5.5.3 Extension Summary

Table 5.16: Deep Brain Stimulation Extension Characteristics

Model/Name	FDA Approval Date	Extensions Enrolled	Extensions Active	Initial Product Performance Events	Median Follow-up Time (Months)	Cumulative Follow-up Time (Months)
37085/37086	March 2009	4,735	2,471	68	37.6	203,009
B34000/B34000M	May 2021	309	290	2	1.2	922

Table 5.17: Deep Brain Stimulation Extension Survival Probability (95% Confidence Intervals)

Model Name	1 Year	2 Years	3 Years	4 Years	5 Years
37085/37086	99.2%	98.9%	98.6%	98.5%	98.1%
	(98.9%, 99.5%)	(98.4%, 99.2%)	(98.2%, 99.0%)	(98.0%, 98.9%)	(97.5%, 98.5%)
B34000/B34000M	—	—	—	—	—
Model Name	6 Years	7 Years	8 Years	9 Years	10 Years
37085/37086	98.1%	97.7%	97.7%	97.7%	97.7%
	(97.5%, 98.5%)	(97.0%, 98.3%)	(97.0%, 98.3%)	(97.0%, 98.3%)	(97.0%, 98.3%)
B34000/B34000M	—	—	—	—	—
Model Name	11 Years				
37085/37086	94.5%	—	—	—	—
	(87.7%, 97.6%)				
B34000/B34000M	—	—	—	—	—