

# Medtronic

## Software compatibility guide



Ensure you have the correct software for the latest LigaSure™ devices

We created this chart of current and recently launched products so you can identify your device, energy platform, and correct software version.

	Order codes	ForceTriad™ energy platform		Valleylab™ FT10 energy platform	
		Software version 3.6 or higher	Software version 3.8 or higher	Software version 1.0 or higher	Software version 1.1 or higher
LigaSure Impact™ nano-coated device	LF4418	✓	✓		✓
LigaSure™ blunt tip nano-coated portfolio	LF1823 LF1837 LF1844	✓	✓		✓
LigaSure™ retractable L-hook device	LF5637 LF5644		✓	✓	✓
LigaSure™ XP Maryland jaw nano-coated portfolio	LXMJ23S LXMJ37S LXMJ44S LXMJ23L LXMJ37L LXMJ44L				✓
LigaSure™ Maryland jaw nano-coated portfolio	LF1923 LF1937 LF1944	✓	✓		✓
LigaSure™ Exact nano-coated dissector	LF2019				✓
LigaSure™ Maryland jaw nano-coated thoracic device	LF1930T				✓



### Questions?

For more information, contact your Medtronic representative or visit **Medtronic.com**

Please ensure your energy platforms have been updated to the compatible software version for your LigaSure™ products.

Refer to your individual LigaSure™ device's instructions for use (IFU) for complete instructions, indications, contraindications, warnings, and precautions.

©2023 Medtronic. Medtronic, Medtronic logo, and Engineering the extraordinary are trademarks of Medtronic. All other brands are trademarks of a Medtronic company. 06/2023 - US-SE-2300101 - [WF#2208504]