



# Software compatibility guide

Ensure you have the correct software for the latest LigaSure™ devices

We created this chart of current and recently launched products so you can identify your device, energy platform, and correct software version.

	Order codes	ForceTriad™ energy platform Software version 3.6 or higher	Valleylab™ FT10 energy platform Software version 1.0 or higher	
LigaSure Impact™ nano-coated device	LF4418	✓	✓	✓
LigaSure™ blunt tip nano-coated portfolio	LF1823 LF1837 LF1844	✓	✓	✓
LigaSure™ retractable L-hook device	LF5637 LF5644		✓	✓
LigaSure™ XP Maryland jaw nano-coated portfolio	LXMJ23S LXMJ37S LXMJ44S LXMJ23L LXMJ37L LXMJ44L			✓
LigaSure™ Maryland jaw nano-coated portfolio	LF1923 LF1937 LF1944	✓	✓	✓
LigaSure™ Exact nano-coated dissector	LF2019			✓
LigaSure™ Maryland jaw nano-coated thoracic device	LF1930T			✓



## Questions?

For more information, contact your Medtronic representative or visit [Medtronic.com](https://Medtronic.com)

Please ensure your energy platforms have been updated to the compatible software version for your LigaSure™ products.

Refer to your individual LigaSure™ device's instructions for use (IFU) for complete instructions, indications, contraindications, warnings, and precautions.